

TCB

**GRANT OF EQUIPMENT
AUTHORIZATION**

TCB

Certification

**Issued Under the Authority of the
Federal Communications Commission**

By:

**Nemko Canada Inc.
303 River Road
Ottawa, Ontario, K1V 1H2
Canada**

Date of Grant: 03/10/2016

Application Dated: 03/09/2016

**VESAG Health Inc
#B202C, 675 US Highway 1
North Brunswick, NJ 08902**

Attention: Rajendra Sadhu , CEO

NOT TRANSFERABLE

EQUIPMENT AUTHORIZATION is hereby issued to the named GRANTEE, and is VALID ONLY for the equipment identified hereon for use under the Commission's Rules and Regulations listed below.

FCC IDENTIFIER: 2ACCFADWAT102C
Name of Grantee: VESAG Health Inc
Equipment Class: PCS Licensed Transmitter worn on body
Notes: MPERSENS

<u>Grant Notes</u>	<u>FCC Rule Parts</u>	<u>Frequency Range (MHZ)</u>	<u>Output Watts</u>	<u>Frequency Tolerance</u>	<u>Emission Designator</u>
	22H	824.7 - 848.31	0.24	0.033 PM	1M27F9W
	24E	1851.25 - 1908.75	0.2	0.023 PM	1M27F9W

Output power listed is ERP for part 22, EIRP for Part 24. The highest reported SAR for head, extremity, and simultaneous transmission exposure conditions are 0.85 W/kg, 1.85 W/kg, and 1.86 W/kg, respectively.

